

## Ultrasound Transducer Condition Report

<b>Date Tested</b> 21/10/2020	<b>Carried out By</b> – I. Rutherford
<b>Manufacturer</b> - Toshiba	<b>Model Number</b> PVT-375BT
<b>Serial Number</b> – TDA1092481	
<b>Transducer Type</b> – Convex	
<b>Date of Manufacture</b> - 2010/09	
<b>System used to Test</b> – Philips iU22	
<b>Cable</b> – Good	
<b>Lens</b> – Good	
<b>Discolouration</b> - Slight	
<b>Overall Condition</b> – Good      © British Medical Auctions	

### Airscans –

